

ETCH GEL GREEN

DIRECTIONS FOR USE

The Etch-Gel is formulated for the convenient preparation of enamel to retain resin based dental materials

FEATURES:

Optimum viscosity for easy extrusion with thixotropic 'non-drip' effect to remain where applied.

Syringe presentation for ease of application.

Bright colour to ensure precise placement is achieved.

INDICATIONS:

Treatment of enamel margins of class I, II, III and V composite restorations to prevent microleakage.

Treatment of enamel surfaces of class IV composite restorations for retention.

Treatment of occlusal surface of posterior teeth prior to application of fissure sealant resin.

Treatment of enamel surfaces of teeth prior to the attachment of orthodontic appliances by means of resin adhesive.

Treatment of the labial surface of teeth prior to application of resin-bonded ceramic or acrylic veneers.

FILLING APPLICATOR SYRINGE: (applies to economy 50ml syringe only)

Remove caps from both syringes; making sure the connector remains on the large syringe. Attach the small applicator syringe to the connector on the large syringe with the piston of the small applicator syringe depressed to the end of the barrel. Ensure that the small applicator syringe is securely attached to the large syringe by twisting the small applicator syringe clockwise until tight. Apply pressure on the piston of the large syringe to deliver a flow of material from the large syringe to the small applicator syringe. If the flow of material shows resistance; pull the piston of the small applicator syringe simultaneously. To detach, retract piston of large syringe slightly to avoid leakage, then remove the small applicator syringe and clean any excess material with a paper towel. Replace appropriate caps to both syringes when not in use.

DIRECTIONS FOR USE:

If the enamel surface has not been freshly exposed by cavity preparation or tooth reduction, cleaning with oil-free prophylaxis paste or pumice and water is recommended. Syringe the etch-gel directly on to the enamel surfaces to be treated. The gel can be spread using the Syringe tip or any other suitable instrument (small ball-ended plunger or brush.) Leave etch gel in place for 30 seconds. Wash and dry using oil-free air. Enamel surface should have a dull frosted appearance. If not, repeat procedure as above.

WARNING:

The Etch-Gel contains 40% phosphoric acid. Corrosive. Causes burns. In case of contact with eyes, rinse immediately with plenty of water and seek medical advice. In case of accident or if you feel unwell, seek medical advice immediately (show label where possible).

PRECAUTIONS:

Secure tip cap immediately after use. Do not force plunger; if blocked, replace tip.

STORAGE:

Store in a cool dry place (5-25°C).

BATCH CODE:



The batch code gives an open date of manufacture in month, year, day format with a numerical suffix to uniquely identify the batch of material. Please quote this batch number in all correspondence.



The expiry date is shown in year, month format. Do not use the product after this date.

SPECIALLY FORMULATED FOR USE IN DENTISTRY

REF: 389X1 12ml Syringe Green Gel



Advanced Healthcare Ltd., Tonbridge, Kent, TN11 8JU, UK

Tel: +44 1892 870500

Fax: +44 1892 870482

Advanced Healthcare Ltd operate a policy of continuous monitoring and improvement of our products. If you have any comments about this product, please contact us at the above address stating the batch number shown on the packaging.

Caution: U.S. Federal Law restricts this device to sale by or on the order of a dental professional.