

ROOT CANAL SEALANT (ZINC OXIDE EUGENOL)

ISO 6876

DIRECTIONS FOR USE

FEATURES:

A paste-paste zinc oxide eugenol cement for use with obturation points. The 2 pastes mix easily to a smooth creamy consistency with long working time and setting time. The cement has moderate flow characteristics, is radiopaque and is low in film thickness (about 6µm).

RECOMMENDED INDICATIONS:

- Permanent obturation of the root canal space with the aid of obturation points.

CONTRA-INDICATIONS

1. Zinc oxide eugenol based sealants in general may produce periapical inflammation in the first few days after placement. Such inflammation should subside naturally as healing progresses.
2. There are no known systemic side-effects to date. However, some people may exhibit sensitivity to eugenol. If any such reactions are experienced, discontinue use of product and seek medical assistance.

APPLICATION:

Base Paste : Catalyst Paste Ratio	2cm : 2cm 0.33g : 0.15g
Mixing Time	20 Seconds
Working Time On Pad (at 23 ± 1°C and 50 ± 5% RH)	30 Minutes minimum
Setting Time In Mouth (at 37 ± 1°C)	30 Minutes minimum

1. Squeeze equal lengths of base paste and catalyst paste on pad and mix thoroughly with flat bladed spatula to creamy consistency of uniform colour.
2. Coat the obturation point with cement.
3. Spread a layer of cement over all inside surfaces of root canal using an appropriate carrier (e.g. endodontic file, gutta percha or paper point).
4. Seat obturation point slowly using normal procedures.

STORAGE:

Store powder in a cool dry, place (5-25°C). Close tubes immediately after use.

BATCH CODE:



The batch code gives an open date of manufacture in month, year, day format with a numerical suffix to uniquely identify the batch of material. Please quote this batch number in all correspondence.



The expiry date is shown in year, month format. Do not use the product after this date.

SPECIALLY FORMULATED FOR USE IN DENTISTRY

REF: 398X1 8g Base Paste / 3.5g Catalyst Paste



Advanced Healthcare Ltd., Tonbridge, Kent, TN11 8JU, UK
Tel: +44 1892 870500 Fax: +44 1892 870482

Advanced Healthcare Ltd operate a policy of continuous monitoring and improvement of our products. If you have any comments about this product, please contact us at the above address stating the batch number shown on the packaging.

Caution: U.S. Federal Law restricts this device to sale by or on the order of a dental professional.