



AHL Generic



LIGHT CURED GLAZE

ISO 6874:2015 Polymer based pit and fissure sealant

DIRECTIONS FOR USE

FEATURES:

A light cure resin glaze designed to protect and enhance glass ionomer restoratives.

INTENDED PURPOSE:

Protect the surface of restorations

INTENDED PATIENT POPULATION:

From child to geriatrics

INTENDED USER:

This product has been formulated for use in dentistry and is intended for use by dental professionals only.

CLINICAL BENEFIT:

To help maintain the integrity of the remaining tooth structure.

INDICATIONS FOR USE:

- Transparent sealant for restorations.
- Transparent fissure sealant

CONTRA-INDICATIONS:

- Do not use on patients with known allergies to resin based materials.
- Do not use with Eugenol containing restorative materials.

Do not use this product for any purposes other than specifically outlined under INDICATIONS in these instructions for use

CONTENTS OF PACK:

Liquid 5ml, instructions for use

PRECAUTIONS AND WARNINGS:

- Do not expose patients or users known to be allergic to this type of material.
- Irritating to oral mucosa, eyes and skin.
- May cause sensitisation by skin contact.
- In case of contact, rinse immediately with plenty of water and seek medical advice.
- Wear suitable gloves and eye/face protection.

PLACEMENT:

- Place the glass ionomer according to manufacturer's instructions.
- When restoration has set, dry surface to be coated do not desiccate
- Dispense a few drops of the coating on to a suitable dish or mixing pad and replace the lid immediately.
- Apply coating with a disposable applicator within 2 minutes of dispensing and cure immediately, using a dental surgery light at 470nm for 20 seconds.
- Re-apply and cure following any subsequent adjustment or polishing.

STORAGE:



Store in a cool, dry place (5-25°C).
Always replace cap immediately after use.

EXPIRY:



The expiry date is shown in year, month format. Do not use the product after this date.

DISPOSAL:

Dispose of the contents and containers in accordance with relevant local and national requirements.

POSSIBLE SIDE EFFECTS / RESIDUAL RISKS:

- This product contains substances that may cause an allergic reaction.

BATCH CODE:



The batch code gives an open date of manufacture in month, year, day format with a numerical suffix to uniquely identify the batch of material. Please quote this batch number in all correspondence.

DEVICE CODES:



126X1

5ml Bottle

COMPOSITION:

Composition	% by weight
Methacrylate resins	80 – 100
Initiators	<1
Stabilisers	<1

AHL operate a policy of continuing surveillance & monitoring of our products. If you experience any incidents relating to the use of this product, please immediately contact us at the below address stating the batch number shown on the packaging. If you experience any serious incident relating to the use of this product, please immediately contact AHL at the below address and the competent authority of the territory you are in..

A summary of safety & clinical performance (SSCP) is available via the EUDAMED database. <https://ec.europa.eu/tools/eudamed>

Caution: U.S. Federal Law restricts this device to sale by or on the order of a dental professional.



Advanced Healthcare Ltd., Tonbridge, Kent, TN11 8JU, UK
Tel: +44 1892 870500 Email: sales@ahl.uk.com



Advena Ltd. Tower Business Centre, 2nd Flr., Tower Street, Swatar, BKR 4013 Malta.

2025-02

AP8074/2

